



NASA Dryden Flight Research Center
PROCUREMENT QUALITY REQUIREMENTS



Q-16 Revision A / Calibration of Procured Devices

The manufacturer's certified calibration report shall be provided with each shipment of a calibrated gauge, instrument, tool or other device sold to the Government.

Each calibration report shall:

- Identify a unique calibration report / tracking number.
- Be traceable to the customer contract / order number.
- Identify the device's name, model number, and when applicable, its serial number.
- Identify the manufacturer's recommended recalibration frequency.
- List the date of initial calibration.
- List the calibration technician's name.
- List the relevant environmental conditions for each parameter calibrated.
- Identify each standard that was used in the device's calibration, including:
 - Each standard's unique identifier, with NIST traceability.
 - The nominal value of each standard, as determined during its most recent calibration.
- Record the value obtained by the device for each standard used in the calibration.