

Q1. What level of percentage of oxygen do you propose to use for these chambers.

A1. NASA plans to use oxygen levels to 100%.

Q2. Is a specific brand of chamber required?

A2. No specific brand is required.